

Manual

RT PCR Mix Probe

Ready-to-use mix for real-time PCR with fluorescent probe.
2x concentrated.

catalog #	size
2008-200P	200 reactions in 25 µl
2008-2000P	2000 reactions in 25 µl

For research use only.

Guarantee

A&A Biotechnology provides guarantee on this product.

The company does not guarantee correct performance of this kit in the event of:

- not adhering to the supplied protocol
- use of not recommended equipment or materials
- use of other reagents than recommended or which are not a component of the product
- use of expired or improperly stored product or its components



Description

RT PCR Mix Probe is optimized ready to use real-time PCR mixture for use with fluorescent probe. Mixture contains all components required for qPCR except DNA template, primers and probes.

The pre-mix formulation saves time and reduces contamination due to a reduced number of pipetting steps required for PCR set up. The mix is optimized for efficient and reproducible reaction.

Contents

	2008-200P	2008-2000P	storage
RT PCR Mix Probe	2 x 1.25 ml	20 x 1.25 ml	-20 °C
ultrapure water	2 x 1.5 ml	20 x 1.5 ml	-20 °C

RT PCR Mix Probe composition

component	amount
<i>Taq</i> DNA polymerase	0.1 U/μl
MgCl ₂	10 mM
dNTPs	0.5 mM of each dNTP
2x reaction buffer	

Notes

- Before use all solutions should be thawed thoroughly on ice, gently mixed by inverting the tube and briefly centrifuged.
- Up to 7x repeated freeze-thaw cycles do not influence the activity of this product.

Example PCR protocol

1. Thaw **all components of the kit** on ice, gently mix by inverting the tubes and briefly centrifuge. Place the tubes on ice again.
2. Place PCR tubes on ice and add:

component	PCR reaction volume		
	10 µl	25 µl	50 µl
RT PCR Mix Probe	5 µl	12.5 µl	25 µl
primer 1 ^{***}	0.1-1 µM*	0.1-1 µM*	0.1-1 µM*
primer 2 ^{***}	0.1-1 µM*	0.1-1 µM*	0.1-1 µM*
probe ^{***}	0.05-0.1 µM**	0.05-0.1 µM**	0.05-0.1 µM**
DNA, cDNA template	10 pg-1 µg	10 pg-1 µg	10 pg-1 µg
ultrapure water	up to 10 µl	up to 25 µl	up to 50 µl

* recommended for standard qPCR

** amount of each probe should be optimized

*** final concentration in reaction mixture

3. Gently vortex the samples and briefly centrifuge to collect all droplets remaining on the tube walls and caps to the bottom of the tube.
4. Place the tubes in the thermocycler and start the PCR programme.

An example amplification profile:

step	temperature	time
initial denaturation	95 °C	2-3 min
25-45 cycles	95 °C	15-30 s
	58-70 °C*	15-60 s**

* depending on the elongation of the probe and primers temperature

** depending on the length of PCR products and/or number of amplicons in the tube

Recommended ROX mixture

HiROX (0.6-1 µl per 50 µl of total reaction volume): Applied Biosystems: 7000, 7300, 7700, 7900HT Fast, StepOne, StepOnePlus.

LowROX (0.6-1 µl per 50 µl of total reaction volume): Applied Biosystems: 7500, Stratagene: Mx3000P, Mx3005P, Mx4000P.



A&A BIOTECHNOLOGY
innovating life science

A&A Biotechnology, ul. Strzelca 40, 80-299 Gdańsk, Poland
phone +48 883 323 761, +48 600 776 268
info@aabiotech.com, www.aabiotech.com

