

### Manual

# Sensitive RT HS-PCR Mix SYBR®

High specificity ready-to-use mix for real-time hot-start PCR with SYBR® Green. 2x concentrated.

catalog#	size
2017-100BM	200 reactions in 25 μl
2017-1000BM	2000 reactions in 25 μl

For research use only.

#### Guarantee

A&A Biotechnology provides guarantee on this product.

- not adhering to the supplied protocol
- $\bullet \qquad \quad \text{use of not recommended equipment or materials}$
- use of other reagents than recommended or which are not a component of the product
- use of expired or improperly stored product or its components

SYBR® is a registered trademark of Molecular Probes Inc.



### Description

Sensitive RT HS-PCR Mix SYBR $^{\otimes}$  is optimized for high specificity ready to use real-time hot-start PCR mixture with SYBR $^{\otimes}$  Green.

Mixture contains all components required for qPCR except DNA template and primers.

Taq DNA polymerase is blocked by monoclonal antibody.

The premix formulation saves time and reduces contamination due to a reduced number of pipetting steps required for PCR set up. The mix is optimized for efficient and reproducible reaction.

#### Contents

	2017-100BM	2017-1000BM	storage
Sensitive SYBR®	2 x 1.25 ml	20 x 1.25 ml	-20 °C, in darkness
ultrapure water	2 x 1.5 ml	20 x 1.5 ml	-20 °C

## Sensitive SYBR® composition

component	amount
modified Taq DNA polymerase	0.1 U/μΙ
MgCI <sub>2</sub>	4 mM
dNTPs	0.5 mM of each dNTP
2x reaction buffer with SYBR® Green	

### **Notes**

- Before use all solutions should be thawed thoroughly on ice, gently mixed by inverting the tube and briefly centrifuged.
- Up to 3x repeated freeze-thaw cycles do not influence the activity of this product.

### **Example PCR protocol**

- 1. Thaw all components of the kit on ice, gently mix by inverting the tubes and briefly centrifuge. Place the tubes on ice again.
- 2. Place PCR tubes on ice and add:

		PCR reaction volume		
component	10 μΙ	25 μΙ	50 μl	
Sensitive SYBR®	5 μΙ	12.5 μΙ	25 μΙ	
primer 1**	0.1-1 μM*	0.1-1 μM*	0.1-1 μM*	
primer 2**	0.1-1 μM*	0.1-1 μM*	0.1-1 μM*	
DNA, cDNA template	10 pg-1 μg	10 pg-1 μg	10 pg-1 μg	
ultrapure water	up to 10 μl	up to 25 μl	up to 50 μl	

<sup>\*</sup> recommended for standard gPCR

- 3. Gently vortex the samples and briefly centrifuge to collect all droplets remaining on the tube walls and caps to the bottom of the tube.
- 4. Place the tubes in the thermocycler and start the PCR programme.

An example amplification profile:

step	temperature	time
initial denaturation	95 ℃	3-5 min
	95℃	15-30 s
25-45 cycles	50-68 °C	30-60 s
	72℃	15-60 s*

<sup>\*</sup> depending on the length of PCR products

#### **Recommended ROX mixture**

HiROX (0.6-1  $\mu$ l per 50  $\mu$ l of total reaction volume): Applied Biosystems: 7000, 7300, 7700, 7900HT Fast, StepOne, StepOnePlus.

**LowROX** (0.6-1  $\mu$ l per 50  $\mu$ l of total reaction volume): Applied Biosystems: 7500, Stratagene: Mx3000P, Mx3005P, Mx4000P.

<sup>\*\*</sup> final concentration in reaction mixture

PCR product melting analysis is recommended.



A&A Biotechnology, ul. Strzelca 40, 80-299 Gdańsk, Poland phone +48 883 323 761, +48 600 776 268 info@aabiot.com, www.aabiot.com

